BACK PLUNGER TYPE DIAL INDICATORS

INSPECTION CERTIFICATE
TRACEABLE TO NIST

handle (included)

2357-5



- Supplied with handle
- With limit pointers
- Optional accessory: contact points (page 102~103)

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading
2357-5*	5mm	0.01mm	15μm	5μm	1mm	0-100

*Supplied with manufacturer inspection certificate traceable to NIST USA

WATERPROOF DIAL INDICATORS

IP54 WATERPROOF

JEWELED BEARING

INSPECTION CERTIFICATE
TRACEABLE TO NIST

2324-10



- Meet DIN878
- IP54 dust/waterproof
- Jeweled bearing

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading	Remark
2324-10F*	10mm	0.01mm	17μm	3μm	1mm	0-100	flat back
2324-10*	10mm	0.01mm	17μm	3μm	1mm	0-100	lug back

*Supplied with manufacturer inspection certificate traceable to NIST USA

DOUBLE FACE DIAL INDICATOR

INSPECTION CERTIFICATE
TRACEABLE TO NIST

front

back

2328-10



- Meet DIN878
- Read from both sides
- Jeweled bearing
- Supplied with limit pointers
- Optional accessory: contact points (page 102~103)

Code	Range	Graduation	Accuracy	Hysteresis	Range/rev	Dial reading
2328-10*	10mm	0.01mm	17μm	3μm	1mm	0-100

*Supplied with manufacturer inspection certificate traceable to NIST USA